

NTATION PAGE

Form Approved
OMB No. 0704-0188

1a. REPORT
Unclassified
2a. SECURITY

AD-A221 984

1b. RESTRICTIVE MARKINGS

3. DISTRIBUTION/AVAILABILITY OF REPORT
Approved for Public Release
Distribution Unlimited

2b. DECLASSIFICATION/DOWNGRADING SCHEDULE

4. PERFORMING ORGANIZATION REPORT NUMBER(S)

5. MONITORING ORGANIZATION REPORT NUMBER(S)

6a. NAME OF PERFORMING ORGANIZATION
North Dakota State University

6b. OFFICE SYMBOL
(if applicable)

7a. NAME OF MONITORING ORGANIZATION
AFOSR/NC

6c. ADDRESS (City, State, and ZIP Code)
Fargo, North Dakota 58105

7b. ADDRESS (City, State, and ZIP Code)
Building 410
Bolling AFB 20332

8a. NAME OF FUNDING/SPONSORING
ORGANIZATION
AFOSR

8b. OFFICE SYMBOL
(if applicable)
NC

9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER
AFOSR
89-0156

8c. ADDRESS (City, State, and ZIP Code)
Building 410
Bolling AFB 20332

10. SOURCE OF FUNDING NUMBERS

PROGRAM ELEMENT NO.	PROJECT NO.	TASK NO.	WORK UNIT ACCESSION NO.
61102F	3842	A2	

11. TITLE (Include Security Classification)

Purchase of a Gas Chromatograph - Mass Spectrometer

12. PERSONAL AUTHOR(S)

Philip Boudjouk

13a. TYPE OF REPORT
Final

13b. TIME COVERED
FROM 88-12-01 TO 89-11-30

14. DATE OF REPORT (Year, Month, Day)
90-04-25

15. PAGE COUNT
3

16. SUPPLEMENTARY NOTATION

17. COSATI CODES

FIELD	GROUP	SUB-GROUP

18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)

Gas chromatograph; Mass spectrometer; GC-MS
Nuclear Instrumentation; Physical Chemistry

19. ABSTRACT (Continue on reverse if necessary and identify by block number)

This report summarizes the final results of the purchase, acquisition and installation of a Hewlett-Packard Model 5988A Gas Chromatograph-Mass Spectrometer.

DTIC
ELECTE
MAY 29 1990
D

20. DISTRIBUTION/AVAILABILITY OF ABSTRACT

☒ UNCLASSIFIED/UNLIMITED ☒ SAME AS RPT. ☐ DTIC USERS

21. ABSTRACT SECURITY CLASSIFICATION

Unclassified

22a. NAME OF RESPONSIBLE INDIVIDUAL

Dr. Fred Hedberg

22b. TELEPHONE (Include Area Code)

(202) 767-4963

22c. OFFICE SYMBOL

NC

Title Page

Final Technical Report
for AFOSR Grant No. 89-0156

Purchase of a Gas Chromatograph - Mass Spectrometer

Philip Boudjouk
Department of Chemistry
North Dakota State University
Fargo, North Dakota 58105

Accession For	
NTIS CRA&I	<input checked="checked" type="checkbox"/>
DTIC TAB	<input type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification	
By	
Distribution /	
Availability Codes	
Dist	Avail and/or Special
A-1	

Abstract

This report summarizes the final results of the purchase, acquisition and installation of a Hewlett-Packard Model 5988A Gas Chromatograph- Mass Spectrometer.

Goals of the Research Project

Funds from the Department of Defense through Grant No. AFOSR 89-0156 were used in combination with matching funds from North Dakota State University to purchase a Hewlett-Packard Model 5988A whose specifications are given below:

HEWLETT PACKARD MODEL 5988A GAS CHROMATOGRAPH - MASS SPECTROMETER

The Basic System:

-MS core with dual EI/CI source, vacuum system, electronics, 220V 60 Hz. Includes capillary direct interface and installation. -5988A MS core includes 2424A CPU with 2.5M byte memory, HP-IB interface, 4 RS232 ports, 3 I/O slots and 9144 CTU in a Design Plus cabinet.

-Gas Chromatograph Mainframe.

Accessories:

-One half inch port into ion source for use with 1/2" probes.

-Direct insertion probe kit for solid samples.

-Capillary Inlet System for Split/Splitless operation. Includes two flow controllers for total flow and septum purge 30 psig gauge, and back pressure regulator.

-HP-IB Data Communications. Includes Inet board, HP-IL to HP-IB converter, and bracket for converter.

-A-Series GC/LC/MS Data System hardware for use with 5970B, 5995C.

-152 Mbyte Disc Drive. Includes HP7958B Disc Drive.

-HP 2235A Rugged Writer Printer, 24 pins, 480 cps (draft) graphics printer with RS-232 and Centronics interface, local printer cable, 16K buffer and paper.

-HP150II Monochrome Graphics Terminal. Includes HP45851A monochrome graphics terminal with cable and two pages alphanumeric memory. Recommended for Aquarius output. HP2397A Color Graphics Terminal. Includes 7 pages alphanumeric memory.

-Interface for 5970B, 5995C, or 5988A. Includes scan box and control HW to interface 5970B with option 410, 5995C with option 410 or 5988A to either option 110 or 120 CPU. Requires 2 CPU I/O slots.

-GC/LC/MS A-Series Software. Includes execute only A-Series operating System and GC/LC/MS applications software with Aquarius II and TOP packages and system installation. Requires 40 M bytes disc space.

-5-day Operators Training Course; which presents the basic operation of an HP Mass Spectrometer with an RTE Data System. Attendees are expected to pay for their travel, food, and lodging RTE System Manager Training Course. A 5-day seminar covering use of an RTE operating system with GC/MS applications software. Requires RTE GC/MS operator training or equivalent.

Conclusion

This instrument is now fully installed and operational. All specifications have been met or exceeded and the instrument is now fully implemented into the research program supported by the Air Force Office of Scientific Research through Grant No. 88-0060 as well as other research projects in the Department of Chemistry.